

MISSOURI DEPARTMENT OF HEALTH STATE PUBLIC HEALTH LABORATORY

JUN -9 2009

DATAMASTER MAINTENANCE REPORT

ISS STOP IT IT OF

Complete this report in duplicate at the time of the is repaired. Send copy to Department of Health; retain	e regular monthly preventive maintenance check, and whenever instrument original in department file.		
204081 127207	DATE OF INSPECTION		
Scott County Jail, Benton, Missouri	TIME OF INSPECTION 5635		
CHECKLIST: Place a check (*) to the left of each it in observed values where determined.) Unchecked ite	em if found to be satisfactory or if operating within established limits. (Write		
☐ DIAGNOSTIC CHECK (PRINTOUT ATTACHED)			
ССОМРИТЕЯ	☐ DETECTOR		
PROGRAM	년 FILTERS		
PHEATERS SAMPLE CHAMBER 50 C	O QUARTZ STANDARD		
LY FLOW DETECTOR	L'CALIBRATION		
PUMP HIGH SPEED	D PRINTER		
☐ INDICATOR LIGHTS			
TIME AND DATE			
SIMULATOR TEMPERATURE (34 °C ± 0.2°C)	34°C		
CALIBRATION CHECK -			
Run three tests using a standard solution. All	three tests must be within \pm 5% of the standard value and must have a nding to the standard solution being used. (PRINTOUT ATTACHED) (USE		
☐ 0.100% STANDARD - MUST READ BETWEEN ☐ 0.040% STANDARD - MUST READ BETWEEN (ONLY ONE STANDARD IS TO BE USED PER M	0.038% and 0.042% INCLUSIVE		
TEST 1 ► . 097 TEST 2	, 097 TEST3 . 098		
PERFORM R.F.I. TEST (PRINTOUT ATTACHED)			
NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)			
REFUSALS () (004) (.0509)	(.1014) 4 (.1519) 2 (Over .19) 2		
List any new parts and describe any alteration or mo and within established limits (use other side if necessa Guth Labs Lot # CS + C Expiration I			
·			
INSPECTING OFFICER			
CONSIGNATION OF THE CONTRACT O	PRINT HAME		
► CV rolling	Greg D. Kenley		
820278 09-29-10	(573) 840-9500		
	(373) 010 2000		



CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Numb. 08340 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography and found to contain 0.1211 percent (w/vol) ethyl alcohol. The expiration date for this lot number is October 15, 2009 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.10 percent BAC.

The alcoho and wat used in this solution were free of test interfering substances.

ed L. Pauley, President GUTH LABORATORIES, INC.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL BAC DATAMASTER SERIAL NUMBER 204081 06/03/09

ARREST TIME: 06:00 SUBJECT NAME:

TEST

00B: 01/01/01

SEX: M

STATE/D.L.: MO/000000

ARRESTING OFFICER:

KENLEY

OFFICER I.D.: 491

TESTING OFFICER:

KENLEY

OFFICER I.D.: 491

PERMIT NUMBER: 820278

EXPIRATION DATE: 09/29/10

MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

CMSU 2208-02

Face This Side Down - This Edge In First

BAC DataMasterEvidence Ticket

MISSOURI STATE HIGHWAY PATROL BAC DATAMASTER SERIAL NUMBER 204081 06/03/09

TESTING OFFICER:
KENLEY/G/D
OFFICER 1.D.: 491
PERMIT NUMBER: 820278
EXPIRATION DATE: 09/29/10
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	06:34
INTERNAL STANDARD	VERIFIED	06:34
EXTERNAL STANDARD	.097	06:35
BLANK TEST	.000	06:36
EXTERNAL STANDARD	.097	Ø6:36
BLANK TEST	.000	06:37
EXTERNAL STANDARD	.098	96:37
BLANK TEST	.000	96:38

N = 3 | SIM. = .1 | AVG. = .0973

Operator Signature 6000

Printed on recycled paper with agri-based lnks

CMSU 2208-02

:):): •

.Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL BAC DATAMASTER SERIAL NUMBER 204081 06/03/09 06:28

--- DIAGNOSTIC CHECK ---

COMPUTER:

OKAY

PROGRAM:

OKAY

HEATERS

SAMPLE CHAMBER:

5⊌с

FLOW DETECTOR:

-OKBY

PUMP

HIGH SPEED:

OKAY

DETECTOR:

OKAY

FILTERS:

OKAY

QUARTZ STANDARD:

OKRY

CALIBRATION:

OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFG HIJKLMNOPQRSTUVWXYZ[\]^_\abcdefghijklmno pqrstuvwxyz{l}+

Operator Signature

60 Kerly

Printed on recycled paper with agri-based links

CMSU 2208-02

State of Missouri DEPARTMENT OF HEALTH



PERMIT TYPE II



GREG D. KENLEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

	•
09/29/08	John J Mathewson
820278	Director of State Public Health Laboratory
Number 09/29/2010	Water the second
Expires	Director, Department of Health
MO 580-0771 (7-88)	Lab. 4 (R7-88)